



VERIFICATION OF COMPLIANCE

Product	Lithium Ion Battery Pack
Model/ Type Ref.	3117615, 3117715, 3117815, 3117915, 3118015, 3118115, 3118215, 3118315 and 3118415. (Mobius 4.0)
Ratings and Principal Characteristics	10.8 VDC, 26100 mAh
Tested According to	UN Manual of Tests and Criteria, Part III, sub-section 38.3
Verification Report No.	3878184BAT-001
Date of Issued	2015.11.17
Applicant	Enovate Medical LLC
Address of Applicant	1152 Park Avenue, Murfreesboro, TN 37129, USA
As shown in the Test Report No. which forms Part of this Verification	3878184BAT-001

Conclusion

Based upon a review of the Test Report(s), the apparatus mentioned above is deemed to comply with the below requirements of:

UN Manual of Tests and Criteria, Fifth Revised Edition, Amendment 2, Part III, sub-section 38.3

Test 1: Altitude Simulation
Test 2: Thermal Test
Test 3: Vibration

Test 4: Shock
Test 5: External Short Circuit
Test 7: Overcharge

Note: This verification is only valid for the product and configuration described and in conjunction with the technical data detailed above.

Reviewed by

Rodney Grimes
Sr. Battery Engineer